



**University of  
Zurich<sup>UZH</sup>**

**Zurich Open Repository and  
Archive**

University of Zurich  
University Library  
Strickhofstrasse 39  
CH-8057 Zurich  
[www.zora.uzh.ch](http://www.zora.uzh.ch)

---

Year: 2018

---

## **Retinopathy of prematurity status during screening: invisible with binocular indirect ophthalmoscopy but established with Optos ultra-wide-field retinal imaging**

Fierz, Fabienne ; Patel, Chetan K

DOI: <https://doi.org/10.1016/j.opthta.2018.11.008>

Posted at the Zurich Open Repository and Archive, University of Zurich

ZORA URL: <https://doi.org/10.5167/uzh-158750>

Journal Article

Accepted Version



The following work is licensed under a Creative Commons: Attribution-NonCommercial-NoDerivatives 4.0 International (CC BY-NC-ND 4.0) License.

Originally published at:


Fierz, Fabienne; Patel, Chetan K (2018). Retinopathy of prematurity status during screening: invisible with binocular indirect ophthalmoscopy but established with Optos ultra-wide-field retinal imaging. *Ophthalmology*:S0161-6420(18)32915-4.

DOI: <https://doi.org/10.1016/j.opthta.2018.11.008>

# Accepted Manuscript

Retinopathy of prematurity status during screening: invisible with binocular indirect ophthalmoscopy but established with Optos ultra-wide-field retinal imaging

Fabienne Fierz, Chetan K. Patel



PII: S0161-6420(18)32915-4

DOI: <https://doi.org/10.1016/j.ophtha.2018.11.008>

Reference: OPHTHA 10551

To appear in: *Ophthalmology*

Received Date: 4 November 2018

Accepted Date: 8 November 2018

Please cite this article as: Fierz F, Patel CK, Retinopathy of prematurity status during screening: invisible with binocular indirect ophthalmoscopy but established with Optos ultra-wide-field retinal imaging, *Ophthalmology* (2018), doi: <https://doi.org/10.1016/j.ophtha.2018.11.008>.

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

Retinopathy of prematurity status during screening: invisible with binocular indirect ophthalmoscopy but established with Optos ultra-wide-field retinal imaging.

Fabienne Fierz<sup>1,2</sup>, Chetan K Patel<sup>\*1</sup>

1. Oxford Eye Hospital, Oxford, UK

2. Department of Ophthalmology, University of Zurich, Zurich, CH-8091 Zurich, Switzerland

\*Corresponding author

Legend

Indirect ophthalmoscopy by an expert failed to visualise the retina because of lens opacity and tunica vasculosa lentis(A), in a baby aged 34 weeks born at 27 weeks weighing 1000grams. The Optos California (Dunfermline, Scotland) has a 0.3mm scanning laser beam that is scattered less than achromatic light and has a virtual focal point behind anterior surface of the lens. The field of view increased as the flying baby technique approximated the eye to the camera(B). Cataract artefact blocked the disc( blue arrow, D) and when the scanning beam entered the eye in an adjacent clear zone(E), retinopathy of prematurity was absent.

